Steve Constant

NOAA

Calibration Report

02/05/2013

Your equipment passed all test and meets factory specifications. This report documents the calibration results.

Equipment	Serial #
6025	07L100056

Final Calibration

Parameter	Result	Min	Max	Comments
Chlorophyll DI Water	0.0			Pass
Chlorophyll 113	113.0			Pass

Repair Tech Comments

Calibrated to factory specifications

FONDRIEST ENVIRONMENTAL, INC.
1415 Research Park Drive - Beavercreek, OH 45432 - Phone: (937) 426-2151 - http://www.Fondriest.com

Steve Constant

NOAA

Calibration Report

02/05/2013

Your equipment passed all test and meets factory specifications. This report documents the calibration results.

Equipment	Serial #
6025	07H101112

Final Calibration

Parameter	Result	Min	Max	Comments
Chlorophyll DI Water	0.0			Pass
Chlorophyll 113	113.0			Pass

Repair Tech Comments

Calibrated to factory specifications

FONDRIEST ENVIRONMENTAL, INC.

1415 Research Park Drive - Beavercreek, OH 45432 - Phone: (937) 426-2151 - http://www.Fondriest.com

Steve Constant

NOAA

Calibration Report 02/05/2013

Your equipment passed all test and meets factory specifications. This report documents the calibration results.

Equipment	Serial #
6150	08D101652

Final Calibration

Parameter	Result	Min	Max	Comments
DO %	95.8			Pass

Repair Tech Comments

Replaced optical DO membrane
Replaced optical DO probe quad seal
Calibrated to factory specifications



Steve Constant

NOAA

Calibration Report 02/05/2013

Your equipment passed all test and meets factory specifications. This report documents the calibration results.

Equipment	Serial #
6150	08E101403

Final Calibration

Parameter	Result	Min	Max	Comments
DO %	95.8			Pass

Repair Tech Comments

Replaced optical DO membrane
Replaced Optical DO quad seal
Calibrated to factory specifications

Steve Constant

NOAA

Calibration Report

02/05/2013

Your equipment passed all test and meets factory specifications. This report documents the calibration results.

Equipment	Serial #
6150	11A100032

Final Calibration

Parameter	Result	Min	Max	Comments
DO %	95.8			Pass

Repair Tech Comments

Replaced optical DO membrane
Replaced Optical DO quad seal
Calibrated to factory specifications



Steve Constant

NOAA

Calibration Report 02/05/2013

Your equipment passed all test and meets factory specifications. This report documents the calibration results.

Equipment	Serial #
6150	09B100770

Final Calibration

Parameter	Result	Min	Max	Comments
DO %	95.8			Pass

Repair Tech Comments

Replaced optical DO membrane
Replaced Optical DO quad seal
Calibrated to factory specifications



Steve Constant

NOAA

Calibration Report 02/05/2013

Your equipment passed all test and meets factory specifications. This report documents the calibration results.

Equipment	Serial #
6025	04K17397AG

Final Calibration

Parameter	Result	Min	Max	Comments
Chlorophyll DI Water	0.0			Pass
Chlorophyll 113	113.0			Pass

Repair Tech Comments

Replaced chlorophyll quad seal
Calibrated to factory specifications

FONDRIEST ENVIRONMENTAL, INC.
1415 Research Park Drive - Beavercreek, OH 45432 - Phone: (937) 426-2151 - http://www.Fondriest.com

Steve Constant

NOAA

Calibration Report 02/05/2013

Your equipment passed all test and meets factory specifications. This report documents the calibration results.

Equipment	Serial #
6025	07K101473

Final Calibration

Parameter	Result	Min	Max	Comments
Chlorophyll DI Water	0.0			Pass
Chlorophyll 113	113.0			Pass

Repair Tech Comments

Replaced chlorophyll probe under warranty
Calibrated to factory specifications

FONDRIEST ENVIRONMENTAL, INC.
1415 Research Park Drive - Beavercreek, OH 45432 - Phone: (937) 426-2151 - http://www.Fondriest.com

Steve Constant

NOAA

Calibration Report

02/05/2013

Your equipment passed all test and meets factory specifications. This report documents the calibration results.

Equipment	Serial #
6025	04D5527AD

Final Calibration

Parameter	Result	Min	Max	Comments
Chlorophyll DI Water	0.0			Pass
Chlorophyll 113	113.0			Pass

Repair Tech Comments

Replaced chlorophyll quad seal
Calibrated to factory specifications

FONDRIEST ENVIRONMENTAL, INC.
1415 Research Park Drive - Beavercreek, OH 45432 - Phone: (937) 426-2151 - http://www.Fondriest.com

Steve Constant

NOAA

Calibration Report 02/05/2013

Your equipment passed all test and meets factory specifications. This report documents the calibration results.

Equipment	Serial #
6025	04F104434

Final Calibration

Parameter	Result	Min	Max	Comments
Chlorophyll DI Water	0.0			Pass
Chlorophyll 113	113.0			Pass

Repair Tech Comments

Replaced chlorophyll quad seal
Calibrated to factory specifications

FONDRIEST ENVIRONMENTAL, INC.
1415 Research Park Drive - Beavercreek, OH 45432 - Phone: (937) 426-2151 - http://www.Fondriest.com

Steve Constant

NOAA

Calibration Report

02/05/2013

Your equipment passed all test and meets factory specifications. This report documents the calibration results.

Equipment	Serial #
6136	07F102090

Final Calibration

Parameter	Result	Min	Max	Comments
Turbidity DI Water	0.0			Pass
Turbidity 126	126.0			Pass

Repair Tech Comments

Replaced turbidity quad seal
Calibrated to factory specifications

FONDRIEST ENVIRONMENTAL, INC.
1415 Research Park Drive - Beavercreek, OH 45432 - Phone: (937) 426-2151 - http://www.Fondriest.com

Steve Constant

NOAA

Calibration Report 02/05/2013

Your equipment passed all test and meets factory specifications. This report documents the calibration results.

Equipment	Serial #
6131	09B100693

Final Calibration

Parameter	Result	Min	Max	Comments
BGA-PC 0	0			Pass
BGA-PC 64120	64120			Pass

Repair Tech Comments

Replaced blue green algae probe quad seal
Calibrated to factory specifications

FONDRIEST ENVIRONMENTAL, INC.
1415 Research Park Drive - Beavercreek, OH 45432 - Phone: (937) 426-2151 - http://www.Fondriest.com

Steve Constant

NOAA

Calibration Report 02/05/2013

Your equipment passed all test and meets factory specifications. This report documents the calibration results.

Equipment	Serial #
6131	12B100951

Final Calibration

Parameter	Result	Min	Max	Comments
BGA-PC 0	0			Pass
BGA-PC 64120	64120			Pass

Repair Tech Comments

Calibrated to factory specifications

Steve Constant

NOAA

Calibration Report 02/05/2013

Your equipment passed all test and meets factory specifications. This report documents the calibration results.

Equipment	Serial #
6136	04K17771AI

Final Calibration

Parameter	Result	Min	Max	Comments
Turbidity DI Water	0.0			Pass
Turbidity 126	126.0			Pass

Repair Tech Comments

Replaced turbidity quad seal
Calibrated to factory specifications

FONDRIEST ENVIRONMENTAL, INC.
1415 Research Park Drive - Beavercreek, OH 45432 - Phone: (937) 426-2151 - http://www.Fondriest.com

Steve Constant

NOAA

Calibration Report 02/05/2013

Your equipment passed all test and meets factory specifications. This report documents the calibration results.

Equipment	Serial #
6131	12C100718

Final Calibration

Parameter	Result	Min	Max	Comments
BGA-PC 0	0			Pass
BGA-PC 64120	64120			Pass

Repair Tech Comments

Calibrated to factory specifications

Steve Constant

NOAA

Calibration Report

02/05/2013

Your equipment passed all test and meets factory specifications. This report documents the calibration results.

Equipment	Serial #
6136	07L100121

Final Calibration

Parameter	Result	Min	Max	Comments
Turbidity DI Water	0.0			Pass
Turbidity 126	126.0			Pass

Repair Tech Comments

Replaced turbidity quad seal
Calibrated to factory specifications

FONDRIEST ENVIRONMENTAL, INC.
1415 Research Park Drive - Beavercreek, OH 45432 - Phone: (937) 426-2151 - http://www.Fondriest.com